

Recall of SYMJEPI (Epinephrine) Injection

Adamis Pharmaceuticals Corporation is voluntarily [recalling certain lots of SYMJEPI \(epinephrine\) Injection](#) 0.15 mg (0.15 mg/0.3 mL) and 0.3 mg (0.3 mg/0.3 mL) Pre-Filled Single-Dose Syringes to the consumer level.

The voluntary recall was issued after two customer complaints were made on three syringes, regarding difficulty dispensing the product.

The batches listed in the table below are being recalled due to the potential clogging of the needle preventing the dispensing of epinephrine. The products are packaged in 2-count Pre-Filled Single-Dose Syringes per carton and were distributed nationwide in the United States and directly to customers and/or medical facilities. The products can be identified by the label containing the US WorldMeds name and logo pictured on the cartons shown [here](#) and [here](#).

Product	Strength	NDC	Lot	Expiration
SYMJEPI (epinephrine) Injection	0.15 mg/0.3 mL	78670-131-02	21101Y	11/30/2022
	0.3 mg/0.3 mL	78670-130-02	21041W	8/31/2022
			21081W	11/30/2022
			21102W	2/28/2023

